



தமிழ்நாடு மாசு கட்டுப்பாடு வாரியம்

From Thiru Md. Nasimuddin, I A S, Chairman (FAC) Tamil Nadu Pollution Control Board, 76, Mount Salai, Guindy, Chennai- 600 032.	To To be Published in TNPCB website and also as per list enclosed
--	---

Letter No: TNPCBd/Labs/ Pur-II/2378-18/2014 date 28.03.2018

Sub: TNPC Board - Labs -Purchase of 1 No of Analytical Balance to
DEL,Perundurai – Limited tender called for- Regarding

Please furnish your sealed (**to be sealed with sealing wax**) and signed Tender for the supply of 1 No of **Analytical Balance** to TNPC Board, DEL, Perundurai as per the specification enclosed in the Annexure on or before **17.04.2018** addressed to the Deputy Director (Labs) ,Tamil Nadu Pollution Control Board, Chennai-32. If the rates quoted are for indigenous item it should be in Rupees and indicating GST, Packing, Forwarding and Delivering Charges. If the rates quoted are for the imported items the price should be quoted in foreign currency including freight, forwarding charges, insurance excluding customs duty. The Board is eligible for concessional customs duty. The Warranty period, offered by you and the delivery period required by you to supply the material should be clearly mentioned in the Tender as per the Technical specification. The actual manufacturers or their authorized agent and stockiest having at least three years experience in these products should only quote. In case of dealers, authorization letter for dealership from the original manufacturer should be enclosed

The sealed cover, super scribed as " **Tender for the supply of "Analytical Balance"** our reference number and due date of tender, should be sent to the following address:-

**THE DEPUTY DIRECTOR (LABS)
TAMIL NADU POLLUTION CONTROL
BOARD (ANNEXE),
76, MOUNT SALAI,
GUINDY,
CHENNAI – 600 032.**

Tender received after the due date will not be considered.

76, மவுண்ட் சாலையி, கிண்டி, சென்னை - 600 032.

பேரன் : 22353134 / 22353135 / 22353136 / 22353137 / 22353138 / 22353139 / 22353140 / 22353141

டெலி கிராம் : 'சிஓஎன்பிஓஎல்' பேக்ஸ் : 044-22353068



தமிழ்நாடு மாசு கட்டுப்பாடு வாரியம்

If the tender is accepted, supply order will be issued to successful bidder under the following terms and conditions stipulated below:-

01	EMD	:	Rs 6000/- (Rupees Six Thousand only) by the Name of Demand Draft in favor of TNPCB, Chennai
01	Payment	:	30 days from the date of effecting full supply
02.	Penalty	:	10% p.a. on the value of goods for every day of delay beyond the time schedule of delivery till the date of completion of supply.
03.	Agreement	:	An agreement has to be concluded with the Tamil Nadu Pollution Control Board in a non-judicial stamp paper to the value of Rs.20/- at your cost within 15 days from the date of the order.
04.	Security Deposit	:	Demand Draft for 5% of the value of the order placed by the Board may be drawn in favor of Tamil Nadu Pollution Control Board, Chennai, from any one of the Nationalised Bank payable at Chennai towards Security Deposit and sent to this office within 7 days from the date of receipt of the order or before effecting the supply, whichever is earlier. This shall be refunded after Three years from the date the relevant supply and installation of the instrument.
05	Delivery Period required	:	21 days from the date of receipt of the supply order.
06.	Place of Delivery	:	DEL, Perundurai


2/4/18
For Chairman


76, மவுண்ட் சாலை, கிண்டி, சென்னை - 600 032.


போன் : 22353134 / 22353135 / 22353136 / 22353137 / 22353138 / 22353139 / 22353140 / 22353141

டெலி கிராம் : 'சி.ஒ.என்பி.ஒ.எல்' பேக்ஸ் : 044-22353068



**TECHNICAL SPECIFICATIONS FOR THE PURCHASE OF ANALYTICAL MICRO BALANCE
CAPACITY 6.1 G, MINIMUM DISPLAY 0.001 MG)**

S.No.	Specifications	Requirement	BIDDER'S RESPONSE: (Bidder shall indicate "Complied" or "Non Complied" as applicable in each row. Technical Literature/published document should be submitted to prove compliance)
01	Capacity	6.1 g	
02	Readability	1 μ g	
03	Repeatability	\pm 1 micrograms	
04	Linearity, micrograms	\pm 4 micrograms	
05	Linearity, 0 – 500 μ g	\pm 2 micrograms	
06	Operating temperature	5° C to 40° C, 85% RH or less	
07	Pan size (mm / in)	\varnothing 30 mm / 1.2 in Std pan \varnothing 50 mm / 1.9 in Std filter	
08	Automatic zero tracking	Standard feature (can be turned off by menu code)	
09	Calibration	Self calibration	
10	Power Source	230 \pm 10 V AC with AC adapter with scale	
11	Standard Equipment	230 \pm 10 V AC Adapter, level indicator, adjustable feet, RS-232, manual	
12	Accessories	All standard accessories including calibration weight of class 'E' in the range	
13	Warranty	Minimum 3 years warranty	


2/4/18
Deputy Director (Labs)
02/04/18